









NITRILE EXAM GLOVES POWDER FREE, STANDARD CUFF

- COATS (an acronym for colloidal oatmeal system) is a patented and unique nitrile glove technology.
- COATS utilises the powerful benefits :-
 - a) of all-natural oats, an FDA-recognised skin protectant
 - b) moisturising barrier between the glove and skin.
- This acts as a preventative :-
 - a) skin irritation
 - b) hydration dermatitis
 - c) eliminates many of the uncomfortable (dry skin & itchy skin)

Product Information

Reorder Order:	CTS38XXS CTS38XS CTS38SS CTS38MM CTS38LL CTS38XL	XX-SMALL X-SMALL SMALL MEDIUM LARGE X-LARGE
Pack Size:	200 gloves per box for XXS to L 180 gloves per box for XL 10 boxes per carton	
Features:	Fingertip textured Powder free Not made with natural rubber latex Chemo drugs tested Lab chemical tested Ambidextrous	
Colour:	White / Blue	
Regulatory Compliance:	ARTG 164563, FDA 510(k) EU 2016/425, REACH, EU 10/2011, EC 1935/2004, MDD 93/42/EEC	
Standards:	ASTM D6319, ASTM D6124, ASTM D5151, ASTM F1671, ASTM D6978, EN 455 part 1, 2 & 3, EN 1186, EN 13130, CENT/TS 14234, EN 420, EN ISO 374-1 (Type B), EN 16523-1, EN 374 part 2,3,4&5, ISO 10993 part 5 & 10, HACCP Certified	
Patents:	Patent 7,691,436; Patent 7,718,240, Patent 7,740,622; Patent 8,075,965; Patent 8,458,818	
Manufacturing Accreditations:	ISO 9001, ISO 13485, EN ISO 13485	

Specifications for Nitrile Examination Glove

Length (mm)	≥230	
Palm Thickness (center of palm)	0.07 ±0.02	
Finger Thickness (13mm ±3mm from tip)	0.09 ±0.02	
Physical Properties	Before Aging	After Aging
Tensile Strength (MPa)	≥18	≥16
Elongation (%)	≥500	≥400
Inspection Levels & AQL	Inspection Level	AQL
Watertightness	G1	1.5
Physical Dimensions	S2	4
Tensile Strength	S2	4
Visual Inspection (Major)	S4	2.5
Visual Inspection (Minor)	S4	4
Particulate Residue	N = 5	≤2mg/glove
Colloidal Oatmeal Content	N = 5	≥5mg/glove